Form

Getinge Product Security Advisory

Document ID: PSA-2026-SEP v02

Publication Date

September 04, 2026
Rastatt, Germany

Overview

Getinge, as a leading medical device manufacturer, is committed to ensuring medical device cybersecurity for our customers. As part of this engagement, Getinge continuously identifies, analyses, and addresses potential vulnerabilities in our products, often in collaboration with customers and researchers. In accordance with Getinge's Coordinated Vulnerability Disclosure process, Getinge is proactively issuing an **End of Guaranteed Support (EOGS)** advisory regarding security for some of its older Getinge Surgical Workplaces product lines. Getinge will no longer provide security updates/patches for these products and will also stop monitoring public reports of incidents about vulnerabilities for these. This notice transfers the security risk from the manufacturer to the operator.

Exploitability

These products are not connected to the hospital’s network or the internet. To exploit potential vulnerabilities, it is necessary to have physical access to these products.

Secure decommissioning

The products do not have any way to electronically store patient information or health data.

Affected Products and solutions

Affected Products and solutions	Remediation
1002.74 Motorised head rest adjustment	---
1131.12 BETASTAR mobile OR table	
1132.2X ALPHASTAR PRO mobile OR table	
1140.14 Table top for paediatric surgery	
1150.02 ALPHAMAQUET OR table column	
1150.15 OR table top for specialist disciplines	
1150.19 Universal table top II	
1150.20 Universal table top	
1150.22 Table top	
1150.23 Urological table top	
1150.25 Table top	
1150.30 Modular universal table top	
1433.01 YUNO OTN mobile OR table	

Affected Products and solutions	Remediation
1955/1975 TRANSMAQUET patient transfer unit T8193000 TRANSMAQUET patient transfer unit	---

Legal Notice

This Product Security Advisory is based on all our findings that we had at the time of publication. However, the facts of the case are being reviewed on an ongoing basis and it is possible that this may result in changed assessments or that assessments contained in this advisory may turn out to be incorrect. We also reserve the right to change or revoke any recommendations. In addition, differences may result from the circumstances of the individual case on site. This information is naturally not available to us and has not been taken into account. Getinge can therefore accept no responsibility that the information presented here is conclusive or comprehensively correct for you. Therefore, please check carefully to what extent deviations can arise for your individual case. If necessary, you will be informed about new findings through the advisories above.

References

Name	Link
---	---

Revision History

Rev.	Date	Description
01	2026-09-04	Initial version
02	2026-10-22	Products 1150.02, 1002.74, T8193000 added