

Use of longer panel cables for Servo ventilators

In view of the situation with the Coronavirus disease COVID-19, we hereby want to provide the following statement regarding the use of a longer panel cable between the patient unit and the Servo ventilators graphical user interface:

- A longer cable would require extensive testing regarding Electromagnetic Interference and compatibility (EMI and EMC).
- We have, at the moment, no validated solution that we can recommend but are looking into different solutions.
- Utilizing cables not provided by Getinge or extending existing cables provided by Getinge may compromise the safety and performance of the ventilator, as described in the User's manual.

Yours sincerely,

Dr. Miray Kärnekull
Medical Director
Maquet Critical Care AB

Jennie Haag
Director Product Management
Maquet Critical Care AB