

## INSTRUCTIONS FOR USE

# Wash-Checks® UH

## Cleaning Monitor for Lumen Instruments

### WC 110

#### Intended Use

The Wash-Checks UH disposable cleaning monitor for lumened instruments is designed to monitor cleaning efficiency in ultrasonic cleaners.

#### Routine Use

1. Open a Wash-Checks H or E holder and place a Wash-Checks UH monitor inside being careful not to touch the test soil. Close the holder by screwing the two sections together.
2. Attach the loaded holder to the connector using the luer lock in the ultrasonic washer or if no connector is available place the holder in the bottom of the basket.
3. Load the ultrasonic washer with a normal load of dirty hollow instruments.
4. Process per normal procedure.
5. Remove the Wash-Checks H or E holder, unscrew the two parts of the holder and remove the Wash-Checks UH monitor.
6. Examine Wash-Checks UH strip for visual cleanliness. If cleaning cycle was effective, blue soil should disappear.



## INSTRUCTIONS FOR USE

# Wash-Checks® UH

## Cleaning Monitor for Lumen Instruments

### WC 110

#### Intended Use

The Wash-Checks UH disposable cleaning monitor for lumened instruments is designed to monitor cleaning efficiency in ultrasonic cleaners.

#### Routine Use

1. Open a Wash-Checks H or E holder and place a Wash-Checks UH monitor inside being careful not to touch the test soil. Close the holder by screwing the two sections together.
2. Attach the loaded holder to the connector using the luer lock in the ultrasonic washer or if no connector is available place the holder in the bottom of the basket.
3. Load the ultrasonic washer with a normal load of dirty hollow instruments.
4. Process per normal procedure.
5. Remove the Wash-Checks H or E holder, unscrew the two parts of the holder and remove the Wash-Checks UH monitor.
6. Examine Wash-Checks UH strip for visual cleanliness. If cleaning cycle was effective, blue soil should disappear.



## INSTRUCTIONS FOR USE

# Wash-Checks® UH

## Cleaning Monitor for Lumen Instruments

### WC 110

#### Intended Use

The Wash-Checks UH disposable cleaning monitor for lumened instruments is designed to monitor cleaning efficiency in ultrasonic cleaners.

#### Routine Use

1. Open a Wash-Checks H or E holder and place a Wash-Checks UH monitor inside being careful not to touch the test soil. Close the holder by screwing the two sections together.
2. Attach the loaded holder to the connector using the luer lock in the ultrasonic washer or if no connector is available place the holder in the bottom of the basket.
3. Load the ultrasonic washer with a normal load of dirty hollow instruments.
4. Process per normal procedure.
5. Remove the Wash-Checks H or E holder, unscrew the two parts of the holder and remove the Wash-Checks UH monitor.
6. Examine Wash-Checks UH strip for visual cleanliness. If cleaning cycle was effective, blue soil should disappear.



## INSTRUCTIONS FOR USE

# Wash-Checks® UH

## Cleaning Monitor for Lumen Instruments

### WC 110

#### Intended Use

The Wash-Checks UH disposable cleaning monitor for lumened instruments is designed to monitor cleaning efficiency in ultrasonic cleaners.

#### Routine Use

1. Open a Wash-Checks H or E holder and place a Wash-Checks UH monitor inside being careful not to touch the test soil. Close the holder by screwing the two sections together.
2. Attach the loaded holder to the connector using the luer lock in the ultrasonic washer or if no connector is available place the holder in the bottom of the basket.
3. Load the ultrasonic washer with a normal load of dirty hollow instruments.
4. Process per normal procedure.
5. Remove the Wash-Checks H or E holder, unscrew the two parts of the holder and remove the Wash-Checks UH monitor.
6. Examine Wash-Checks UH strip for visual cleanliness. If cleaning cycle was effective, blue soil should disappear.



### Interpretation of Results

Examine the aluminum strip for presence of any test soil remaining. The blue soil should disappear in an efficient ultrasonic wash cycle. The presence of a large amount of soil remaining on the strip means a serious deficiency in ultrasonic washer efficiency. A small amount of soil remaining indicates that conditions are slightly deficient.

### Storage and Handling

Store Wash-Checks UH in the box at 40° - 86°F (4° - 30°C). Keep away from disinfectants and fumes. Remove only the strips needed for the test at hand and re-seal the vial. The expiry date is printed on the product packaging.

---

DF-203 Rev. E



SteriTec Products Mfg. Co., Inc.  
Member of the Getinge Group  
74 Inverness Drive East  
Englewood, CO 80112 USA

[www.getinge.com/steritec](http://www.getinge.com/steritec)

### Interpretation of Results

Examine the aluminum strip for presence of any test soil remaining. The blue soil should disappear in an efficient ultrasonic wash cycle. The presence of a large amount of soil remaining on the strip means a serious deficiency in ultrasonic washer efficiency. A small amount of soil remaining indicates that conditions are slightly deficient.

### Storage and Handling

Store Wash-Checks UH in the box at 40° - 86°F (4° - 30°C). Keep away from disinfectants and fumes. Remove only the strips needed for the test at hand and re-seal the vial. The expiry date is printed on the product packaging.

---

DF-203 Rev. E



SteriTec Products Mfg. Co., Inc.  
Member of the Getinge Group  
74 Inverness Drive East  
Englewood, CO 80112 USA

[www.getinge.com/steritec](http://www.getinge.com/steritec)

### Interpretation of Results

Examine the aluminum strip for presence of any test soil remaining. The blue soil should disappear in an efficient ultrasonic wash cycle. The presence of a large amount of soil remaining on the strip means a serious deficiency in ultrasonic washer efficiency. A small amount of soil remaining indicates that conditions are slightly deficient.

### Storage and Handling

Store Wash-Checks UH in the box at 40° - 86°F (4° - 30°C). Keep away from disinfectants and fumes. Remove only the strips needed for the test at hand and re-seal the vial. The expiry date is printed on the product packaging.

---

DF-203 Rev. E



SteriTec Products Mfg. Co., Inc.  
Member of the Getinge Group  
74 Inverness Drive East  
Englewood, CO 80112 USA

[www.getinge.com/steritec](http://www.getinge.com/steritec)

### Interpretation of Results

Examine the aluminum strip for presence of any test soil remaining. The blue soil should disappear in an efficient ultrasonic wash cycle. The presence of a large amount of soil remaining on the strip means a serious deficiency in ultrasonic washer efficiency. A small amount of soil remaining indicates that conditions are slightly deficient.

### Storage and Handling

Store Wash-Checks UH in the box at 40° - 86°F (4° - 30°C). Keep away from disinfectants and fumes. Remove only the strips needed for the test at hand and re-seal the vial. The expiry date is printed on the product packaging.

---

DF-203 Rev. E



SteriTec Products Mfg. Co., Inc.  
Member of the Getinge Group  
74 Inverness Drive East  
Englewood, CO 80112 USA

[www.getinge.com/steritec](http://www.getinge.com/steritec)