

GETINGE CLEAN

Product Specification

Enzymatic Detergent

PRODUCT DESCRIPTION

A neutral enzymatic detergent for use in washer disinfectors, ultrasonic baths and manual cleaning of medical devices. Suitable for use on water resistant surfaces such as stainless steel, ceramics, glass, plastics and soft metals.

DIRECTION OF USE

Dosing range 2-10ml/l.

Cleaning temperature (machine) 35-60°C.

Cleaning temperature range (manual/ultrasonic) 30-50°C.

Do not use on polycarbonate material.

Check materials compatibility before use.

Ensure complete rinsing after use.

Do not mix with other chemicals.

INGREDIENTS

Protease, lipase, amylase enzymes, surfactants, sequestering agents and corrosion inhibitors.

TECHNICAL DATA

Specific gravity (20°C): 1,047 (typical)

pH as supplied (typical): 7.9

pH at 5ml/water (typical): 7.7

Viscosity: <10 cps

STORAGE INFORMATION

Store in a cool, dry place protected from frost and away from acids and strong oxidizing agents.

Store upright in original containers.

Optimum storage temperature 5-25°C.

SAFETY AND HANDLING

Refer to Material Safety Data Sheet.

CE MARK

Enzymatic Detergent conforms to the European directive 93/42/EC, Annex 1 concerning medical devices.

ENVIRONMENTAL AWARENESS

Biodegradable formulation and recyclable packaging.



All Goods



Detergent



Getinge's range of detergents have been developed, tested and validated on Getinge washers to consistently meet ISO 15883 standards for cleaning efficacy of medical devices.




GETINGE CLEAN

Product Specification

Enzymatic Detergent

WATER HARDNESS AND DETERGENT DOSING



Note:

Water hardness is only one factor that influences the amount of detergent to be dosed.

Other factors to be considered include:

- Duration of cleaning cycle
- Temperature
- Soil type, amount and condition e.g. dry
- Materials compatibility

SERVICE

For technical support in regards to appropriate detergent dosing and optimization of your cleaning processes, please contact your local Getinge representative.

GETINGE MANUAL & AUTOMATED DETERGENT RANGE



	Aluminum	Anodised aluminium	Titanium	Non ferrous metals	Matt chrome plated	Stainless steel	Glass	Ceramic	Rubber	Plastic
■ GETINGE CLEAN HEAVY SOIL DETERGENT			M		●	●	●	●	●	M
■ GETINGE CLEAN UNIVERSAL DETERGENT	M	M	M	M	●	●	●	●	●	M
■ GETINGE CLEAN ENZYMATIC DETERGENT	●	●	●	●	●	●	●	●	●	M
■ GETINGE CLEAN NEUTRALIZER	M	M	●	●	●	●	●	●	M	M
■ GETINGE CLEAN RINSE AID	●	●	●	●	●	●	●	●	M	M
■ GETINGE CLEAN INSTRUMENT LUBRICANT	●	●	●	●	●	●				
■ GETINGE CLEAN PRE TREATMENT FOAM	●	●	●	●	●	●	●	●	●	●

This chart is only a guideline and information on the reprocessing of medical devices from the manufacturer according to EN ISO 17664 must always be followed.

M=Always refer to instrument manufacturer's guidance.

ORDERING NUMBER

5 l	10 l	25 l	205 l
503437005	503437105	503437205	503437305

GETINGE

GETINGE GROUP

Getinge Disinfection AB, Ljungadalsgatan 11
 Box 1505, SE-351 15 Växjö, Sweden
 Phone: +46 (0)10 335 98 00
 Web: <http://www.getinge.com>
 E-mail: info@getinge.com